

Issued by NMI Certin B.V.,  
designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 17 of Directive 2014/32/EU, after having established that the Measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer Flow Meter Group B.V.  
Innovatieweg 32  
7007 CD Doetinchem  
The Netherlands

Measuring instrument **A Rotary Displacement Gas Meter**

Type	:	FMR FMR-Dual
Manufacturer's mark or name	:	Flow Meter Group B.V.
Destined for the measurement of	:	Gas volume
Accuracy class	:	Class 1,0 or 1,5 (see tables under 1.2.1 of the description)
Environment classes	:	M1 / E2
Temperature range	:	-40 °C / +70 °C

Further properties are described in the annexes:

- Description T10372 revision 12;
- Documentation folder T10372-9.

Valid until 30 August 2031

Remarks This revision replaces the earlier versions, except for its documentation folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**  
30 August 2021

## Certification Board

**NMI Certin B.V.**  
Thijsseweg 11  
2629 JA Delft  
The Netherlands  
T +31 88 636 2332  
[certin@nmi.nl](mailto:certin@nmi.nl)  
[www.nmi.nl](http://www.nmi.nl)

This document is issued under the provision that no liability is accepted and that the manufacturer shall indemnify third-party liability.

The designation of NMI Certin B.V. as Notified Body can be verified at <http://ec.europa.eu/growth/tools-databases/nando/>

Reproduction of the complete document only is permitted.

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.